



Decreasing fibrinogen levels are the **first sign of post-partum hemorrhage**, the leading cause of death during childbirth, worldwide

INTRODUCING NEW  
**qLabs<sup>®</sup> FIB**  
Fibrinogen Monitoring System 


A POINT-OF-CARE ANALYSER FOR PPH MANAGEMENT THAT MEASURES FIBRINOGEN LEVELS WITHIN 1 TO 10 MINUTES



The **lower** the level, the **faster the time-to-result**, helping clinicians to give the right product at the right moment to treat post-partum hemorrhage.




From a single drop of blood, the novel POC device delivers robust, lab comparable results, with integrated hematocrit adjustment



### Designed for ease and speed

- Dry reagent strips stored at room temperature for immediate use
- Individually packed in humidity-resistant pouch
- Lyophilized QC
- Touchscreen handheld rapid testing platform with option of eStation II docking system
- Built-in QR code scanner as well as communications port for printers and computer networks



**qLabs® FIB** test shows very strong correlation with Stago STA®-Liquid Fib test. The correlation coefficient is 0.985 with an intercept of 0.034 and slope of 1.006 (n=98)

### Features and specs

#### Quantification of functional fibrinogen

Clot-based fibrinogen measurement correlated to Clauss fibrinogen method

#### Sample Type

Venous whole blood in standard coagulation citrated tube (blue cap, 3.2% (0.109 M) sodium citrate)

#### Sample Volume

15 µL

#### Reportable Range

1.0 - 4.0 g/L (Reference range: 2.0 - 4.0 g/L)

#### Precision

CV ≤7.0%

#### Hematocrit Range

20% to 60%

#### Shelf Life at Room Temperature

24 months

Designation	Cat. Nr.	Packaging
qLabs® FIB meter	Q-3 Plus	Unit
eStation II docking station	MBI92	Unit
qLabs® FIB Test strips	QS-18 Pro	24 strips/box
qLabs® FIB Controls level 1	QS-18-CLN	4 x 1 mL
qLabs® FIB Controls level 2	QS-18-CLP	4 x 1 mL